

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	
m	m	mm	mm	(Uncertainty)		m	m	mm	mm	(Uncertainty)	
0 a 0.01	0.00987	0.130	± 0.2	±	4.3E-01 mm						
0 a 0.02	0.01986	0.140	± 0.2	±	4.3E-01 mm						
0 a 0.03	0.02983	0.170	± 0.2	±	4.3E-01 mm						
0 a 0.04	0.03984	0.160	± 0.2	±	4.3E-01 mm						