

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 0.1	0.09975	0.250	± 0.6	±	4.3E-01 mm						
0 a 0.2	0.19990	0.100	± 0.6	±	4.3E-01 mm						
0 a 0.4	0.39984	0.160	± 0.7	±	4.3E-01 mm						
0 a 0.6	0.59979	0.210	± 0.8	±	4.3E-01 mm						
0 a 0.8	0.79986	0.140	± 0.9	±	4.3E-01 mm						
0 a 1	0.99991	0.090	± 1	±	4.3E-01 mm						
0 a 1.2	1.19987	0.130	± 1	±	4.8E-01 mm						
0 a 1.4	1.40007	-0.070	± 1.1	±	4.8E-01 mm						
0 a 1.6	1.60023	-0.230	± 1.2	±	4.8E-01 mm						
0 a 1.8	1.80019	-0.190	± 1.3	±	4.8E-01 mm						
0 a 2	2.00010	-0.100	± 1.4	±	4.8E-01 mm						
0 a 2.34	2.33977	0.230	± 1.5	±	4.8E-01 mm						