

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.107 °C	10.0 °C	-0.107 °C	± 0.4 °C	± 4.0E-01 °C
40 °C	40.692 °C	40.2 °C	-0.492 °C	± 0.4 °C	± 4.0E-01 °C
25 %HR	25.253 %HR	28.0 %HR	2.747 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.509 %HR	52.0 %HR	1.491 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.356 %HR	74.0 %HR	-1.356 %HR	± 5.0 %HR	± 1.9E+00 %HR