

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.107 °C	10.0 °C	-0.107 °C	± 1.0 °C	± 4.0E-01 °C
40 °C	40.692 °C	39.7 °C	-0.992 °C	± 1.0 °C	± 4.0E-01 °C
25 %HR	25.253 %HR	23.0 %HR	-2.253 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.509 %HR	47.0 %HR	-3.509 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.356 %HR	70.0 %HR	-5.356 %HR	± 5.0 %HR	± 1.9E+00 %HR