

Patrón <i>(Standard)</i>	UBP <i>(UT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1800.452 s	1800.45 s	-0.002 s	-0.0001%	± 0.010%	± 9.2E-02 s
3599.359 s	3599.35 s	-0.009 s	-0.0003%	± 0.010%	± 9.2E-02 s
7199.486 s	7199.48 s	-0.006 s	-0.0001%	± 0.010%	± 1.1E-01 s
17999.953 s	17999.95 s	-0.003 s	0.0000%	± 0.010%	± 1.5E-01 s
32399.641 s	32399.64 s	-0.001 s	0.0000%	± 0.010%	± 2.6E-01 s