

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.282 °C	15.8 °C	0.518 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.179 °C	25.6 °C	0.421 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	51.923 °C	52.3 °C	0.377 °C	± 1.0 °C	± 1.1E+00 °C
25 %HR	21.040 %HR	22.7 %HR	1.660 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.693 %HR	52.6 %HR	1.907 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	72.625 %HR	74.6 %HR	1.975 %HR	± 5.0 %HR	± 1.9E+00 %HR