

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 15 °C                      | 15.282 °C                  | 15.5 °C            | 0.218 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 25 °C                      | 25.179 °C                  | 25.5 °C            | 0.321 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 50 °C                      | 51.923 °C                  | 52.2 °C            | 0.277 °C               | ± 0.5 °C                      | ± 1.1E+00 °C                         |
| 25 %HR                     | 21.040 %HR                 | 23 %HR             | 1.460 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 50.693 %HR                 | 52 %HR             | 1.407 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 75 %HR                     | 72.625 %HR                 | 74 %HR             | 1.175 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |