

Certificado No. (Certificate #):		59952	Equipo (Instrument):		Reloj Palpador
Fecha de Calibración (Calibration Date):		2025-08-14	Marca (Brand):		Mitutoyo
Patrón (Standard)		UBP (UUT)	Error (Error)	Exactitud (Accuracy)	
0.050 in		0.050 in	0.000 in	±	5.8E-04 in
0.100 in		0.100 in	0.000 in	±	5.8E-04 in
0.200 in		0.200 in	0.000 in	±	5.9E-04 in
0.300 in		0.300 in	0.000 in	±	6.1E-04 in
0.400 in		0.400 in	0.000 in	±	6.4E-04 in
0.500 in		0.500 in	0.000 in	±	6.8E-04 in
0.600 in		0.600 in	0.000 in	±	7.1E-04 in
0.800 in		0.800 in	0.000 in	±	8.0E-04 in
1.000 in		1.000 in	0.000 in	±	9.1E-04 in