

Certificado No. (Certificate #):		59954	Equipo (Instrument):		Reloj Palpador
Fecha de Calibración (Calibration Date):		2025-08-14	Marca (Brand):		Mitutoyo
Patrón (Standard)		UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0.050	in	0.050	in	0.000	in ± 5.8E-04
0.100	in	0.100	in	0.000	in ± 5.8E-04
0.200	in	0.200	in	0.000	in ± 5.9E-04
0.300	in	0.300	in	0.000	in ± 6.1E-04
0.400	in	0.400	in	0.000	in ± 6.4E-04
0.500	in	0.500	in	0.000	in ± 6.8E-04
0.600	in	0.600	in	0.000	in ± 7.1E-04
0.800	in	0.800	in	0.000	in ± 8.0E-04
1.000	in	1.000	in	0.000	in ± 9.1E-04