

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.664 °C	30.5 °C	-0.164 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.478 %HR	37.5 %HR	2.022 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.509 %HR	48.2 %HR	-2.309 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	68.0 %HR	-2.243 %HR	± 3.0 %HR	± 1.9E+00 %HR