

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
51.00 mmHg	50 mmHg	-1 mmHg	± 5 mmHg	±	6.6E-01 mmHg
101.00 mmHg	100 mmHg	-1 mmHg	± 5 mmHg	±	6.5E-01 mmHg
152.00 mmHg	150 mmHg	-2 mmHg	± 5 mmHg	±	6.6E-01 mmHg
201.00 mmHg	200 mmHg	-1 mmHg	± 5 mmHg	±	6.8E-01 mmHg
252.00 mmHg	250 mmHg	-2 mmHg	± 5 mmHg	±	7.5E-01 mmHg
298.00 mmHg	300 mmHg	2 mmHg	± 5 mmHg	±	7.5E-01 mmHg