

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
1.00 mm	0.98 mm	-20.00 µm	± 30 µm	±	5.8E+00 µm
1.50 mm	1.48 mm	-20.00 µm	± 30 µm	±	5.8E+00 µm
3.00 mm	2.98 mm	-20.00 µm	± 30 µm	±	5.8E+00 µm
5.00 mm	4.99 mm	-10.00 µm	± 30 µm	±	5.8E+00 µm
10.00 mm	9.97 mm	-30.00 µm	± 30 µm	±	5.8E+00 µm
15.00 mm	14.97 mm	-30.00 µm	± 30 µm	±	5.8E+00 µm
20.00 mm	19.97 mm	-30.00 µm	± 30 µm	±	5.8E+00 µm
50.00 mm	49.97 mm	-30.00 µm	± 30 µm	±	5.8E+00 µm