

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.6	0.595000	5.270	± 1.6	± 4.3E-01 mm					
0 a 0.7	0.695000	5.270	± 1.7	± 4.3E-01 mm					
0 a 0.8	0.795000	5.270	± 1.8	± 4.3E-01 mm					
0 a 0.9	0.895000	5.280	± 1.9	± 4.3E-01 mm					
0 a 1	0.998000	2.280	± 2	± 4.3E-01 mm					
0 a 1.2	1.198000	2.290	± 2.1	± 4.8E-01 mm					
0 a 1.4	1.398000	2.300	± 2.3	± 4.8E-01 mm					
0 a 1.6	1.598000	2.310	± 2.4	± 4.8E-01 mm					
0 a 1.8	1.797000	3.320	± 2.6	± 4.8E-01 mm					
0 a 2	1.997000	3.330	± 2.8	± 4.8E-01 mm					