

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.522 °C	16.1 °C	0.578 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.650 °C	25.2 °C	-0.450 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	50.067 °C	49.7 °C	-0.367 °C	± 1.0 °C	± 1.1E+00 °C
25 %HR	25.458 %HR	23.0 %HR	-2.458 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.967 %HR	47.0 %HR	-2.967 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.356 %HR	73.0 %HR	-2.356 %HR	± 5.0 %HR	± 1.9E+00 %HR