

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.522 °C	15.2 °C	-0.322 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.650 °C	24.9 °C	-0.750 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	50.067 °C	49.4 °C	-0.667 °C	± 1.0 °C	± 1.1E+00 °C
25 %HR	25.458 %HR	23.0 %HR	-2.458 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.967 %HR	48.0 %HR	-1.967 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.356 %HR	72.0 %HR	-3.356 %HR	± 5.0 %HR	± 1.9E+00 %HR