

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.522 °C	15.6 °C	0.078 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.650 °C	25.6 °C	-0.050 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	50.067 °C	50.9 °C	0.833 °C	± 1.0 °C	± 1.1E+00 °C
25 %HR	25.458 %HR	24.0 %HR	-1.458 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.967 %HR	49.0 %HR	-0.967 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.356 %HR	77.0 %HR	1.644 %HR	± 5.0 %HR	± 1.9E+00 %HR