

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
10.20 μm	10.30 μm	0.10 μm	± 2.5 μm	±	5.8E-02 μm
12.10 μm	12.2 μm	0.10 μm	± 2.5 μm	±	5.8E-02 μm
24.10 μm	24.2 μm	0.10 μm	± 2.5 μm	±	5.8E-02 μm
50.60 μm	50.6 μm	0.00 μm	± 2.5 μm	±	5.8E-02 μm