

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
20 °C	20.436 °C	20.5 °C	0.064 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.363 °C	30.3 °C	-0.063 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	40.182 %HR	42.0 %HR	1.818 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	74.3 %HR	3.852 %HR	± 5.0 %HR	± 1.9E+00 %HR