

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
20 °C	20.436 °C	21.5 °C	1.064 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.363 °C	31.6 °C	1.237 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	40.182 %HR	67.0 %HR	26.818 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	95.2 %HR	24.752 %HR	± 5.0 %HR	± 1.9E+00 %HR