

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
20 °C	20.436 °C	20.7 °C	0.264 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.363 °C	30.5 °C	0.137 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	40.182 %HR	43.2 %HR	3.018 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	72.3 %HR	1.852 %HR	± 5.0 %HR	± 1.9E+00 %HR