

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
20 °C	20.436 °C	20.4 °C	-0.036 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.363 °C	30.4 °C	0.037 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	40.182 %HR	44.8 %HR	4.618 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	73.3 %HR	2.852 %HR	± 5.0 %HR	± 1.9E+00 %HR