

Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	E.M.P (<i>Tolerance</i>)	Incertidumbre (<i>Uncertainty</i>)	
1.00 mm	1.00 mm	0.00 μm	± 4 μm	±	5.8E+00 μm
1.50 mm	1.50 mm	0.00 μm	± 4 μm	±	5.8E+00 μm
2.00 mm	2.00 mm	0.00 μm	± 4 μm	±	5.8E+00 μm
5.00 mm	5.00 mm	0.00 μm	± 4 μm	±	5.8E+00 μm
10.00 mm	10.00 mm	0.00 μm	± 4 μm	±	5.8E+00 μm
15.00 mm	15.00 mm	0.00 μm	± 4 μm	±	5.8E+00 μm
20.00 mm	20.00 mm	0.00 μm	± 4 μm	±	5.8E+00 μm
25.00 mm	25.00 mm	0.00 μm	± 4 μm	±	5.8E+00 μm
30.00 mm	30.00 mm	0.00 μm	± 4 μm	±	5.8E+00 μm
35.00 mm	35.00 mm	0.00 μm	± 4 μm	±	5.8E+00 μm
39.00 mm	39.00 mm	0.00 μm	± 4 μm	±	5.8E+00 μm