

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-20 °C	-20.296 °C	-20.9 °C	-0.604 °C	± 1.0 °C	± 4.0E-01 °C
0 °C	-0.131 °C	-0.9 °C	-0.769 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	20.145 °C	19.6 °C	-0.545 °C	± 1.0 °C	± 4.0E-01 °C
50 %HR	49.231 %HR	54 %HR	4.769 %HR	± 5.0 %HR	± 1.9E+00 %HR