

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-20 °C	-20.296 °C	-19.8 °C	0.496 °C	± 0.5 °C	± 4.0E-01 °C
0 °C	-0.131 °C	0.3 °C	0.431 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.145 °C	20.4 °C	0.255 °C	± 0.5 °C	± 4.0E-01 °C
50 %HR	49.231 %HR	53 %HR	3.769 %HR	± 5.0 %HR	± 1.9E+00 %HR