

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.524 °C	30.0 °C	-0.524 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.059 %HR	29.3 %HR	-0.759 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.079 %HR	51.2 %HR	1.121 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	69.0 %HR	-1.448 %HR	± 3.0 %HR	± 1.9E+00 %HR