

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 1	0.99916	0.840	± 1	±	4.3E-01 mm						
0 a 2	1.99925	0.750	± 1.4	±	4.8E-01 mm						
0 a 3	2.99948	0.520	± 1.8	±	4.8E-01 mm						
0 a 4	3.99956	0.440	± 2.2	±	4.8E-01 mm						
0 a 5	4.99962	0.380	± 2.6	±	4.8E-01 mm						
0 a 6	5.99969	0.310	± 3	±	3.6E-01 mm						
0 a 7	6.99982	0.180	± 3.4	±	3.6E-01 mm						
0 a 8	7.99996	0.040	± 3.8	±	3.6E-01 mm						
0 a 9	9.00011	-0.110	± 4.2	±	3.6E-01 mm						
0 a 10	10.00020	-0.200	± 4.6	±	3.6E-01 mm						