

Certificado No. (Certificate #):		60098		Equipo (Instrument):		Medidor de Espesores	
Fecha de Calibración (Calibration Date):		2025-08-20		Marca (Brand):		Defelsko	
Patrón (Standard)		UBP (UUT)		Error (Error)		E.M.P (Tolerance)	
1.0000 mm		1.04 mm		40 µm		± 5.8E+00 µm	
1.5000 mm		1.55 mm		45 µm		± 5.8E+00 µm	
2.0000 mm		2.05 mm		52 µm		± 5.8E+00 µm	
5.0000 mm		5.05 mm		50 µm		± 5.8E+00 µm	
10.0000 mm		9.99 mm		-15 µm		± 5.8E+00 µm	
20.0000 mm		19.89 mm		-108 µm		± 5.8E+00 µm	
30.0000 mm		29.89 mm		-113 µm		± 5.8E+00 µm	
50.0000 mm		49.79 mm		-215 µm		± 5.8E+00 µm	
100.0000 mm		97.90 mm		-2100 µm		± 5.9E+00 µm	