

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.381 °C	0.0 °C	-0.381 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.993 °C	4.9 °C	-0.093 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.007 °C	9.7 °C	-0.307 °C	± 0.5 °C	± 4.0E-01 °C