

| Nominal <i>(Nominal)</i> | Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | Exactitud <i>(Accuracy)</i> | Incertidumbre <i>(Uncertainty)</i> |
|--------------------------|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|
| 0 °C                     | 0.381 °C                 | 0.8 °C           | 0.419 °C             | ± 0.5 °C                    | ± 4.0E-01 °C                       |
| 5 °C                     | 4.993 °C                 | 5.2 °C           | 0.207 °C             | ± 0.5 °C                    | ± 4.0E-01 °C                       |
| 10 °C                    | 10.007 °C                | 10.4 °C          | 0.393 °C             | ± 0.5 °C                    | ± 4.0E-01 °C                       |