

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.381 °C	0.4 °C	0.019 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.993 °C	4.9 °C	-0.093 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.007 °C	10.1 °C	0.093 °C	± 0.5 °C	± 4.0E-01 °C