

Nominal <i>(Nominal)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>
0 °C	0.381 °C	0.5 °C	0.119 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.993 °C	5.2 °C	0.207 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.007 °C	10.1 °C	0.093 °C	± 0.5 °C	± 4.0E-01 °C