

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.564 °C	30.2 °C	-0.364 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.059 %HR	27.4 %HR	-2.659 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.775 %HR	46.2 %HR	-4.575 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	74.8 %HR	4.557 %HR	± 5.0 %HR	± 1.9E+00 %HR