

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.564 °C	29.3 °C	-1.264 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.059 %HR	34.0 %HR	3.941 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.775 %HR	54.0 %HR	3.225 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	72.0 %HR	1.757 %HR	± 5.0 %HR	± 1.9E+00 %HR