

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.280 °C	0.0 °C	-0.280 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.536 °C	20.2 °C	-0.336 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.578 °C	35.3 °C	-0.278 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.059 %HR	33.8 %HR	3.741 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	50.775 %HR	51.3 %HR	0.525 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.264 %HR	77.4 %HR	-2.864 %HR	± 3.0 %HR	± 1.9E+00 %HR