

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.280 °C	0.5 °C	0.220 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.536 °C	20.5 °C	-0.036 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.578 °C	35.5 °C	-0.078 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.059 %HR	33.0 %HR	2.941 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	50.775 %HR	52.4 %HR	1.625 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.264 %HR	79.8 %HR	-0.464 %HR	± 3.0 %HR	± 1.9E+00 %HR