

Tipo de Medida (<i>Measurement Type</i>)	Patrón (<i>Standard</i>)	UBP (<i>UT</i>)	Error (<i>Error</i>)	E.M.P (<i>Tolerance</i>)	Incertidumbre (<i>Uncertainty</i>)
Profundidad / (Depth)	0 mm	0.11 mm	10 μm	± 30 μm	± 6.1E+00 μm
Profundidad / (Depth)	1 mm	0.51 mm	10 μm	± 30 μm	± 5.8E+00 μm
Profundidad / (Depth)	1 mm	1.01 mm	10 μm	± 30 μm	± 6.1E+00 μm
Profundidad / (Depth)	2 mm	2.02 mm	20 μm	± 30 μm	± 6.1E+00 μm
Profundidad / (Depth)	5 mm	5.02 mm	20 μm	± 30 μm	± 6.1E+00 μm
Profundidad / (Depth)	10 mm	10.02 mm	20 μm	± 30 μm	± 6.1E+00 μm
Profundidad / (Depth)	15 mm	15.02 mm	20 μm	± 30 μm	± 6.1E+00 μm
Profundidad / (Depth)	20 mm	20.02 mm	20 μm	± 30 μm	± 6.1E+00 μm
Profundidad / (Depth)	25 mm	25.03 mm	30 μm	± 30 μm	± 6.1E+00 μm