| Patrón (Standard) | UBP (UUT) | Error (Error) | E.M.P (Tolerance)              | Incertidumbre (Uncertainty) |
|-------------------|-----------|---------------|--------------------------------|-----------------------------|
| 2.40 in           | 2.4000 in | 0 in          | ± 0.0003 "                     | ± 7.0E-02 in                |
| 2.45 in           | 2.4500 in | 0 in          | $\pm~0.0003~^{\prime\prime}$   | ± 7.6E-02 in                |
| 2.50 in           | 2.5000 in | 0 in          | $\pm~0.0003~^{\prime\prime}$   | ± 9.4E-02 in                |
| 2.55 in           | 2.5500 in | 0 in          | $\pm~0.0003$ $^{\prime\prime}$ | ± 7.6E-02 in                |
| 2.60 in           | 2.6000 in | 0 in          | $\pm~0.0003~"$                 | ± 9.4E-02 in                |
| 2.65 in           | 2.6500 in | 0 in          | $\pm~0.0003~"$                 | ± 6.9E-02 in                |
| 2.70 in           | 2.7000 in | 0 in          | $\pm~0.0003~"$                 | ± 1.1E-01 in                |
| 2.75 in           | 2.7500 in | 0 in          | ± 0.0003 "                     | ± 1.3E-01 in                |
| 2.80 in           | 2.8000 in | 0 in          | ± 0.0003 "                     | ± 1.4E-01 in                |