Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
2.40 in	2.4000 in	0 in	± 0.0003 "	± 6.00E-05 in
2.45 in	2.4500 in	0 in	$\pm~0.0003$ "	± 6.00E-05 in
2.50 in	2.5000 in	0 in	± 0.0003 "	± 6.00E-05 in
2.55 in	2.5500 in	0 in	± 0.0003 "	± 6.00E-05 in
2.60 in	2.6000 in	0 in	± 0.0003 "	± 6.00E-05 in
2.65 in	2.6500 in	0 in	± 0.0003 "	± 6.00E-05 in
2.70 in	2.7000 in	0 in	± 0.0003 "	± 6.00E-05 in
2.75 in	2.7500 in	0 in	± 0.0003 "	± 6.00E-05 in
2.80 in	2.8000 in	0 in	± 0.0003 "	± 6.00E-05 in