

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
12 °C	12.554 °C	13.0 °C	0.446 °C	± 0.5 °C	± 4.0E-01 °C
27 °C	27.054 °C	27.4 °C	0.346 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	42.296 °C	42.0 °C	-0.296 °C	± 0.5 °C	± 4.0E-01 °C
52 %HR	52.554 %HR	54.3 %HR	1.746 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	72.186 %HR	72.7 %HR	0.514 %HR	± 3.0 %HR	± 1.9E+00 %HR
# 92 %HR	92.227 %HR	89.7 %HR	-2.527 %HR	± 3.0 %HR	± 1.9E+00 %HR