

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.408 °C	10.7 °C	0.292 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.336 °C	20.7 °C	0.364 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.213 °C	30.7 °C	0.487 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.391 °C	40.7 °C	0.309 °C	± 0.5 °C	± 4.0E-01 °C