

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	2.336 °C	2.3 °C	-0.036 °C	± 1.0 °C	± 4.0E-01 °C
5 °C	5.094 °C	4.9 °C	-0.194 °C	± 1.0 °C	± 4.0E-01 °C
8 °C	8.242 °C	8.0 °C	-0.242 °C	± 1.0 °C	± 4.0E-01 °C