

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.408 °C	10.5 °C	0.062 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	20.336 °C	20.3 °C	-0.036 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.213 °C	29.9 °C	-0.313 °C	± 1.0 °C	± 4.0E-01 °C
40 °C	40.391 °C	39.9 °C	-0.491 °C	± 1.0 °C	± 4.0E-01 °C