

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.564 °C	30.9 °C	0.336 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.059 %HR	31.0 %HR	0.941 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.775 %HR	52.0 %HR	1.225 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	68.0 %HR	-2.243 %HR	± 5.0 %HR	± 1.9E+00 %HR