

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.564 °C	30.7 °C	0.136 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.059 %HR	27.0 %HR	-3.059 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.775 %HR	45.8 %HR	-4.975 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	64.8 %HR	-5.443 %HR	± 5.0 %HR	± 1.9E+00 %HR