

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.564 °C	29.6 °C	-0.964 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.059 %HR	27.0 %HR	-3.059 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.775 %HR	46.0 %HR	-4.775 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	65.0 %HR	-5.243 %HR	± 5.0 %HR	± 1.9E+00 %HR