

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.213 °C	30.4 °C	0.187 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.059 %HR	28.1 %HR	-1.959 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.243 %HR	49.8 %HR	-0.443 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	68.9 %HR	-1.548 %HR	± 3.0 %HR	± 1.9E+00 %HR