

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.213 °C	30.3 °C	0.087 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.059 %HR	28.6 %HR	-1.459 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.243 %HR	50.5 %HR	0.257 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	69.2 %HR	-1.248 %HR	± 3.0 %HR	± 1.9E+00 %HR