

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	14.931 °C	15.5 °C	0.569 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.125 °C	25.1 °C	-0.025 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	31.884 °C	32.3 °C	0.416 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.261 %HR	31.9 %HR	2.639 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.292 %HR	70.4 %HR	1.108 %HR	± 3.0 %HR	± 1.9E+00 %HR