

Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)	
50.30 mmHg	50 mmHg	-0.30 mmHg	± 5 mmHg	±	6.6E-01 mmHg
100.10 mmHg	100 mmHg	-0.10 mmHg	± 5 mmHg	±	6.5E-01 mmHg
150.30 mmHg	150 mmHg	-0.30 mmHg	± 5 mmHg	±	6.6E-01 mmHg
200.20 mmHg	200 mmHg	-0.20 mmHg	± 5 mmHg	±	6.8E-01 mmHg
249.70 mmHg	250 mmHg	0.30 mmHg	± 5 mmHg	±	7.5E-01 mmHg
299.80 mmHg	300 mmHg	0.20 mmHg	± 5 mmHg	±	7.5E-01 mmHg