

Patrón (Standard)		UBP (UIT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
7 Hz	0.400 g-pk	0.38 g-pk	-0.02 g-pk	± 0.060 g-pk	±	0.011 g-pk
10 Hz	0.800 g-pk	0.83 g-pk	0.03 g-pk	± 0.12 g-pk	±	0.017 g-pk
50 Hz	1.000 g-pk	1.03 g-pk	0.03 g-pk	± 0.15 g-pk	±	0.021 g-pk
100 Hz	1.000 g-pk	1.03 g-pk	0.03 g-pk	± 0.15 g-pk	±	0.021 g-pk
500 Hz	1.000 g-pk	1.03 g-pk	0.03 g-pk	± 0.15 g-pk	±	0.021 g-pk
1000 Hz	1.000 g-pk	1.05 g-pk	0.05 g-pk	± 0.15 g-pk	±	0.022 g-pk